
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006628
8:24 16/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:08
STERILIZING 4:13
136.4 C 220kPa 4:13
136.1 C 220kPa 4:43
135.8 C 218kPa 5:13
135.9 C 220kPa 5:43
136.3 C 220kPa 6:13
136.4 C 221kPa 6:43
136.4 C 221kPa 7:13
136.4 C 222kPa 7:43
136.4 C 222kPa 7:43
VENTING 7:43
AIR DRYING 8:33
CYCLE COMPLETE 8:40

Digital Signature #
5A7A5FDF07E8B203
