

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008391  
7:41 10/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:55  
PRESSURIZING 1:59  
STERILIZING 3:08  
136.4 C 222kPa 3:08  
136.4 C 222kPa 3:38  
136.3 C 222kPa 4:08  
136.1 C 217kPa 4:38  
136.4 C 221kPa 5:08  
136.2 C 219kPa 5:38  
136.5 C 221kPa 6:08  
136.3 C 222kPa 6:38  
136.3 C 222kPa 6:38  
VENTING 6:39  
AIR DRYING 7:30  
CYCLE COMPLETE 37:30

-----  
Digital Signature #  
C16D9E46B80EF304  
-----