

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001939  
9:31 06/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:38  
PRESSURIZING 3:39  
STERILIZING 4:58  
136.4 C 220kPa 4:58  
136.3 C 223kPa 5:28  
136.4 C 218kPa 5:58  
136.4 C 221kPa 6:28  
136.2 C 224kPa 6:58  
136.4 C 223kPa 7:28  
136.2 C 222kPa 7:58  
136.5 C 224kPa 8:28  
136.5 C 224kPa 8:28  
VENTING 8:28  
AIR DRYING 9:23  
CYCLE COMPLETE 9:24

-----  
Digital Signature #  
4DC94A4BA8F1B03B  
-----