

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2488

Name: RITTER 1  
Sterilizer ID: V2178198  
10/25/2022 03:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.3 F 0.0  
2:00 135.7 F 0.2  
4:00 147.4 F 1.6  
6:00 169.8 F 4.9  
8:00 190.4 F 8.3  
10:00 208.2 F 12.4  
12:00 220.9 F 5.7  
14:00 233.3 F 9.9  
16:00 236.2 F 9.5  
18:00 241.9 F 11.7  
20:00 238.9 F 9.8  
22:00 252.0 F 16.2  
24:00 263.1 F 22.6  
26:00 272.5 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.7 F 28.9  
2:00 271.7 F 28.2  
3:00 271.8 F 28.6  
4:00 272.5 F 28.8

Min 271.2 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.7 F 0.0  
5:00 198.3 F 0.0  
10:00 191.0 F 0.0  
15:00 193.1 F 0.0  
20:00 194.8 F 0.0  
25:00 209.8 F 0.0  
30:00 226.1 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 26:19  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:04:38  
CYCLE COMPLETE

10/25/2022 04:43 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_