
Midmark M11 - v1.0.0
Total Cycles: 1128

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2021 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.9 F 0.0
2:00 110.1 F 0.0
4:00 130.6 F 0.1
6:00 161.7 F 0.3
8:00 186.3 F 3.2
10:00 199.3 F 6.6
12:00 214.5 F 10.6
14:00 227.6 F 6.6
16:00 233.4 F 9.5
18:00 240.3 F 10.3
20:00 243.3 F 12.0
22:00 240.1 F 10.3
24:00 252.8 F 16.6
26:00 263.7 F 23.0
28:00 272.2 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.4 F 28.9
2:00 272.0 F 28.4
3:00 271.1 F 28.1
4:00 272.3 F 29.0

Min 271.0 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.3 F 0.0
5:00 226.3 F 0.0
10:00 225.9 F 0.0
15:00 227.2 F 0.0
20:00 232.1 F 0.0
25:00 231.4 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 28:14
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:06:20
CYCLE COMPLETE

08/30/2021 10:16 AM
ACCEPTED: X
REJECTED:
Operator: U02