
Midmark M11 - v1.0.5
Total Cycles: 2423

Name: RITTER2
Sterilizer ID: V2313228
05/26/2023 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.2 F 0.0
2:00 144.0 F 2.1
4:00 155.4 F 5.2
6:00 183.9 F 9.2
8:00 203.4 F 10.3
10:00 218.7 F 7.4
12:00 230.6 F 12.0
14:00 232.0 F 8.6
16:00 241.0 F 11.5
18:00 244.2 F 12.8
20:00 255.9 F 19.0
22:00 266.5 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.6 F 29.8
2:00 271.8 F 29.1
3:00 271.2 F 29.2
4:00 272.6 F 29.8

Min 271.0 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 204.3 F 0.0
10:00 208.2 F 0.0
15:00 207.7 F 0.0
20:00 208.9 F 0.0
25:00 210.0 F 0.0
30:00 211.5 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 23:42
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:02:17
CYCLE COMPLETE

05/26/2023 04:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____