
Midmark M11 - v1.0.5
Total Cycles: 4358

Name: RITTER2
Sterilizer ID: V2313228
02/19/2025 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.0 F 0.0
2:00 157.3 F 1.8
4:00 157.7 F 4.9
6:00 187.6 F 8.8
8:00 197.0 F 13.2
10:00 220.8 F 6.9
12:00 231.6 F 11.2
14:00 230.6 F 7.7
16:00 241.3 F 13.2
18:00 241.9 F 11.8
20:00 253.3 F 17.7
22:00 263.6 F 23.9
24:00 272.3 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 272.6 F 30.6
2:00 272.1 F 30.5
3:00 272.1 F 30.3
4:00 272.2 F 30.1

Min 270.9 F 29.3
Max 273.7 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 205.2 F 0.0
10:00 222.8 F 0.0
15:00 210.4 F 0.0
20:00 200.7 F 0.0
25:00 198.0 F 0.0
30:00 195.8 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 24:15
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:03:15
CYCLE COMPLETE

02/19/2025 02:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____