

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3636

Name: RITTER2  
Sterilizer ID: V2313228  
05/22/2024 02:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.1 F 0.0  
2:00 145.7 F 1.9  
4:00 154.7 F 5.0  
6:00 181.1 F 9.0  
8:00 202.6 F 11.8  
10:00 219.7 F 7.3  
12:00 230.4 F 12.4  
14:00 233.2 F 9.6  
16:00 236.1 F 9.4  
18:00 246.5 F 14.4  
20:00 258.3 F 21.0  
22:00 269.1 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.5  
1:00 272.6 F 30.3  
2:00 272.2 F 29.7  
3:00 271.2 F 29.5  
4:00 272.6 F 30.4

Min 271.2 F 29.4  
Max 273.1 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.9 F 0.0  
5:00 207.5 F 0.0  
10:00 207.7 F 0.0  
15:00 213.2 F 0.0  
20:00 214.5 F 0.0  
25:00 211.9 F 0.0  
30:00 211.4 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 23:10  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:01:35  
CYCLE COMPLETE

05/22/2024 03:35 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_