
Midmark M11 - v1.0.5
Total Cycles: 2234

Name: RITTER2
Sterilizer ID: V2313228
04/11/2023 03:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.5 F 0.0
2:00 189.3 F 4.0
4:00 208.2 F 9.3
6:00 227.3 F 8.1
8:00 235.7 F 11.2
10:00 235.0 F 9.2
12:00 237.5 F 9.6
14:00 249.2 F 15.4
16:00 261.6 F 22.5
18:00 272.7 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.6 F 29.7
2:00 271.6 F 29.4
3:00 272.7 F 29.8
4:00 271.7 F 29.1

Min 271.5 F 29.1
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 229.4 F 0.0
10:00 224.8 F 0.0
15:00 225.2 F 0.0
20:00 222.2 F 0.0
25:00 221.8 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 2:28
HEATING: 18:24
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

04/11/2023 04:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____