
Midmark M11 - v1.0.5
Total Cycles: 3709

Name: RITTER2
Sterilizer ID: V2313228
06/18/2024 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 71.5 F 0.0
4:00 78.9 F 0.4
6:00 116.0 F 2.2
8:00 144.0 F 4.7
10:00 162.0 F 8.1
12:00 191.3 F 12.2
14:00 217.1 F 4.5
16:00 226.8 F 8.0
18:00 238.4 F 11.8
20:00 229.8 F 7.7
22:00 241.2 F 13.0
24:00 240.2 F 10.8
26:00 252.1 F 16.8
28:00 262.2 F 22.8
30:00 272.2 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 271.8 F 29.7
2:00 272.2 F 29.9
3:00 272.5 F 29.9
4:00 272.7 F 29.8

Min 271.4 F 29.3
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
5:00 222.5 F 0.0
10:00 196.0 F 0.0
15:00 197.4 F 0.0
20:00 195.8 F 0.0
25:00 191.4 F 0.0
30:00 199.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:06
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:09:15
CYCLE COMPLETE

06/18/2024 09:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____