
Midmark M11 - v1.0.0
Total Cycles: 4420

Name: RITTER 1
Sterilizer ID: V2178198
04/02/2024 09:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 88.1 F 0.2
4:00 81.6 F 0.2
6:00 110.4 F 0.6
8:00 148.1 F 3.2
10:00 166.8 F 5.6
12:00 181.3 F 8.5
14:00 197.2 F 11.7
16:00 208.5 F 2.7
18:00 221.1 F 6.0
20:00 230.4 F 9.3
22:00 239.9 F 13.0
24:00 236.5 F 9.4
26:00 241.1 F 10.5
28:00 245.4 F 12.7
30:00 256.3 F 18.5
32:00 266.2 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.2 F 28.4
2:00 272.5 F 28.9
3:00 271.5 F 28.2
4:00 272.0 F 28.5

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.1 F 0.0
5:00 210.4 F 0.0
10:00 209.9 F 0.0
15:00 197.9 F 0.0
20:00 192.0 F 0.0
25:00 191.1 F 0.0
30:00 203.0 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 33:42
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:12:11
CYCLE COMPLETE

04/02/2024 10:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____