
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001907
9:18 29/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:51
STERILIZING 4:06
136.0 C 221kPa 4:06
136.4 C 225kPa 4:36
136.3 C 222kPa 5:06
136.2 C 224kPa 5:36
136.3 C 223kPa 6:06
136.3 C 222kPa 6:36
136.1 C 223kPa 7:06
136.3 C 224kPa 7:36
136.3 C 224kPa 7:36
VENTING 7:36
AIR DRYING 8:23
CYCLE COMPLETE 38:23

Digital Signature #
766FE7F692646712
