
Midmark M11 - v1.0.5
Total Cycles: 385

Name: RITTER2
Sterilizer ID: V2313228
09/22/2021 01:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.0 F 0.0
2:00 182.0 F 3.5
4:00 192.2 F 9.4
6:00 216.8 F 6.4
8:00 230.4 F 11.4
10:00 235.2 F 9.8
12:00 236.8 F 9.0
14:00 250.7 F 16.4
16:00 263.3 F 23.7
18:00 271.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.4 F 29.6
2:00 272.6 F 30.0
3:00 272.1 F 29.9
4:00 272.1 F 29.5

Min 271.6 F 29.3
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 226.2 F 0.0
10:00 233.5 F 0.0
15:00 226.2 F 0.0
20:00 224.0 F 0.0
25:00 225.1 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 18:01
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:57:15
CYCLE COMPLETE

09/22/2021 02:03 PM
ACCEPTED: X
REJECTED:
Operator: U01