
Midmark M11 - v1.0.0
Total Cycles: 3117

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2023 03:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.8 F 0.0
2:00 206.0 F 2.6
4:00 203.3 F 7.9
6:00 236.4 F 12.9
8:00 233.3 F 9.0
10:00 239.4 F 10.3
12:00 243.2 F 12.1
14:00 240.5 F 10.6
16:00 254.0 F 17.1
18:00 265.2 F 23.8
20:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 28.6
1:00 272.5 F 28.8
2:00 272.5 F 28.9
3:00 272.6 F 29.2
4:00 272.0 F 28.9

Min 271.8 F 28.5
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 225.5 F 0.0
10:00 236.7 F 0.0
15:00 232.1 F 0.0
20:00 226.6 F 0.0
25:00 228.1 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

04/14/2023 04:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____