
Midmark M11 - v1.0.5
Total Cycles: 1814

Name: RITTER2
Sterilizer ID: V2313228
12/06/2022 03:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.6 F 0.0
2:00 148.8 F 0.0
4:00 166.0 F 2.3
6:00 197.1 F 7.3
8:00 218.1 F 12.7
10:00 228.9 F 8.1
12:00 240.3 F 13.5
14:00 239.2 F 11.3
16:00 238.5 F 10.0
18:00 252.0 F 16.8
20:00 263.8 F 23.8
22:00 271.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.5 F 29.8
2:00 271.9 F 29.1
3:00 271.7 F 29.3
4:00 272.5 F 29.7

Min 271.4 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 219.7 F 0.0
5:00 212.0 F 0.0
10:00 224.7 F 0.0
15:00 231.8 F 0.0
20:00 227.9 F 0.0
25:00 227.5 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:04
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

12/06/2022 04:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____