
Midmark M11 - v1.0.0
Total Cycles: 4785

Name: RITTER 1
Sterilizer ID: V2178198
08/02/2024 01:23 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.9 F 0.0
2:00 149.4 F 3.2
4:00 184.6 F 9.1
6:00 199.9 F 5.7
8:00 223.0 F 7.9
10:00 235.2 F 11.4
12:00 239.5 F 11.1
14:00 237.3 F 9.1
16:00 251.9 F 16.3
18:00 238.3 F 9.5
20:00 255.0 F 17.7
22:00 268.0 F 25.6
24:00 238.5 F 9.7
26:00 258.6 F 19.7
28:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.0
1:00 272.9 F 28.7
2:00 272.0 F 28.0
3:00 272.5 F 28.7

Min 271.7 F 27.9
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.9 F 0.0
5:00 190.7 F 0.0
10:00 200.5 F 0.0
15:00 207.5 F 0.0
20:00 209.8 F 0.0
25:00 215.2 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 28:37
STERILIZING: 3:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:05:33
CYCLE COMPLETE

08/02/2024 02:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____