

-----  
Midmark M11 - v1.0.5  
Total Cycles: 708

Name: RITTER2  
Sterilizer ID: V2313228  
01/19/2022 03:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.0 F 0.0  
2:00 172.4 F 4.0  
4:00 196.7 F 9.2  
6:00 219.1 F 7.6  
8:00 230.8 F 11.0  
10:00 233.6 F 9.2  
12:00 237.3 F 9.5  
14:00 249.4 F 15.7  
16:00 262.1 F 22.9  
18:00 272.6 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.3 F 29.7  
2:00 271.6 F 29.6  
3:00 272.4 F 29.7  
4:00 271.3 F 29.3  
  
Min 271.3 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
5:00 215.2 F 0.0  
10:00 212.5 F 0.0  
15:00 216.3 F 0.0  
20:00 225.1 F 0.0  
25:00 225.3 F 0.0  
30:00 221.8 F 0.0  
  
Drying Complete

FILLING: 1:48  
HEATING: 18:14  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:13  
CYCLE COMPLETE

01/19/2022 04:17 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_