

-----  
Midmark M11 - v1.0.5  
Total Cycles: 174

Name: RITTER2  
Sterilizer ID: V2313228  
07/07/2021 11:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.1 F 0.0  
2:00 158.0 F 3.0  
4:00 175.4 F 6.6  
6:00 195.3 F 11.4  
8:00 218.5 F 5.5  
10:00 230.3 F 10.7  
12:00 232.3 F 7.6  
14:00 243.8 F 13.0  
16:00 244.0 F 12.6  
18:00 256.1 F 19.1  
20:00 266.8 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.3 F 29.5  
2:00 271.9 F 29.7  
3:00 272.2 F 29.4  
4:00 271.9 F 29.8

Min 271.5 F 29.1  
Max 272.5 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.3 F 0.0  
5:00 199.4 F 0.0  
10:00 214.2 F 0.0  
15:00 216.5 F 0.0  
20:00 215.9 F 0.0  
25:00 215.1 F 0.0  
30:00 213.7 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 21:28  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:52  
CYCLE COMPLETE

07/07/2021 12:53 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02