
Midmark M11 - v1.0.5
Total Cycles: 1204

Name: RITTER2
Sterilizer ID: V2313228
06/21/2022 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.6 F 0.0
2:00 135.1 F 1.4
4:00 145.6 F 4.4
6:00 180.7 F 8.4
8:00 197.6 F 13.5
10:00 224.2 F 8.0
12:00 233.0 F 12.8
14:00 234.9 F 9.3
16:00 238.7 F 10.1
18:00 244.9 F 13.1
20:00 256.7 F 19.4
22:00 266.8 F 25.9
24:00 273.5 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 272.7 F 29.6
2:00 272.3 F 30.0
3:00 271.9 F 30.0
4:00 272.1 F 29.7

Min 271.4 F 29.5
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.5 F 0.0
5:00 220.2 F 0.0
10:00 228.3 F 0.0
15:00 206.1 F 0.0
20:00 197.0 F 0.0
25:00 209.6 F 0.0
30:00 204.0 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 24:16
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

06/21/2022 09:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____