
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001601
11:46 16/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:22
STERILIZING 3:24
136.0 C 220kPa 3:24
136.4 C 224kPa 3:54
136.4 C 225kPa 4:24
136.4 C 224kPa 4:54
136.1 C 223kPa 5:24
136.1 C 222kPa 5:54
136.2 C 222kPa 6:24
136.2 C 222kPa 6:54
136.2 C 222kPa 6:54
VENTING 6:54
AIR DRYING 7:46
CYCLE COMPLETE 7:56

Digital Signature #
755D5033778D0C91
