

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2814

Name: RITTER 1  
Sterilizer ID: V2178198  
01/24/2023 02:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
2:00 187.1 F 2.3  
4:00 196.7 F 6.8  
6:00 213.8 F 12.2  
8:00 225.2 F 7.4  
10:00 238.0 F 12.9  
12:00 238.9 F 10.5  
14:00 236.2 F 8.9  
16:00 251.5 F 16.1  
18:00 263.7 F 23.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.6 F 29.0  
2:00 271.5 F 28.3  
3:00 272.6 F 29.2  
4:00 272.1 F 28.6  
  
Min 271.5 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 226.1 F 0.0  
10:00 234.3 F 0.0  
15:00 228.6 F 0.0  
20:00 223.7 F 0.0  
25:00 222.6 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 20:00  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:34  
CYCLE COMPLETE

01/24/2023 03:23 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_