
Midmark M11 - v1.0.5
Total Cycles: 1892

Name: RITTER2
Sterilizer ID: V2313228
01/04/2023 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.3 F 0.0
2:00 174.0 F 3.0
4:00 179.8 F 7.0
6:00 202.4 F 12.1
8:00 218.0 F 6.6
10:00 233.7 F 11.8
12:00 235.5 F 9.5
14:00 238.4 F 9.2
16:00 248.3 F 15.0
18:00 260.6 F 22.0
20:00 271.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.1 F 29.8
2:00 271.3 F 29.3
3:00 272.3 F 29.9
4:00 271.5 F 29.1

Min 271.2 F 29.1
Max 272.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.5 F 0.0
5:00 231.4 F 0.0
10:00 228.1 F 0.0
15:00 230.8 F 0.0
20:00 219.4 F 0.0
25:00 220.1 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 20:33
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:58:55
CYCLE COMPLETE

01/04/2023 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____