
Midmark M11 - v1.0.5
Total Cycles: 2004

Name: RITTER2
Sterilizer ID: V2313228
02/02/2023 05:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.4 F 0.0
2:00 199.2 F 4.6
4:00 210.6 F 11.1
6:00 224.7 F 7.3
8:00 239.7 F 12.9
10:00 241.6 F 12.2
12:00 242.3 F 11.8
14:00 256.7 F 19.3
16:00 269.0 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.0 F 28.9
2:00 272.5 F 29.3
3:00 272.4 F 29.6
4:00 271.7 F 28.8

Min 271.5 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 220.9 F 0.0
10:00 238.0 F 0.0
15:00 231.7 F 0.0
20:00 223.2 F 0.0
25:00 222.2 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 2:35
HEATING: 16:58
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:56:37
CYCLE COMPLETE

02/02/2023 06:36 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____