
Midmark M11 - v1.0.0
Total Cycles: 4070

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2023 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.8 F 0.0
2:00 89.1 F 0.5
4:00 101.2 F 1.8
6:00 138.4 F 4.0
8:00 166.8 F 7.2
10:00 180.0 F 10.8
12:00 195.1 F 4.7
14:00 209.8 F 5.1
16:00 219.9 F 8.5
18:00 230.0 F 12.3
20:00 231.5 F 7.9
22:00 242.4 F 12.8
24:00 240.2 F 10.6
26:00 251.5 F 16.2
28:00 261.5 F 21.9
30:00 271.1 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.3 F 29.2
2:00 272.2 F 28.8
3:00 271.4 F 28.3
4:00 271.6 F 28.7

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 212.3 F 0.0
10:00 210.1 F 0.0
15:00 208.6 F 0.0
20:00 205.7 F 0.0
25:00 217.2 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:49
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:09:22
CYCLE COMPLETE

11/30/2023 12:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____