
Midmark M11 - v1.0.0
Total Cycles: 1530

Name: RITTER 1
Sterilizer ID: V2178198
01/17/2022 08:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 78.2 F 0.0
4:00 95.4 F 0.0
6:00 119.0 F 0.3
8:00 151.6 F 2.9
10:00 174.8 F 6.2
12:00 189.9 F 9.5
14:00 204.6 F 13.2
16:00 216.5 F 5.7
18:00 226.5 F 9.4
20:00 236.6 F 13.5
22:00 236.3 F 10.0
24:00 236.6 F 8.9
26:00 246.4 F 13.7
28:00 257.5 F 19.8
30:00 267.8 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.2 F 28.9
2:00 272.7 F 29.3
3:00 271.8 F 28.9
4:00 272.1 F 28.9

Min 271.6 F 28.6
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 214.4 F 0.0
10:00 227.6 F 0.0
15:00 224.8 F 0.0
20:00 225.4 F 0.0
25:00 224.3 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:19
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:27
CYCLE COMPLETE

01/17/2022 09:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____