
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005401
10:07 21/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 2:56
STERILIZING 4:01
136.4 C 223kPa 4:01
136.4 C 224kPa 4:31
136.1 C 223kPa 5:01
136.4 C 225kPa 5:31
136.2 C 222kPa 6:01
136.4 C 224kPa 6:31
136.1 C 218kPa 7:01
136.4 C 220kPa 7:31
136.4 C 220kPa 7:31
VENTING 7:31
AIR DRYING 8:21
CYCLE COMPLETE 8:46

Digital Signature #
C0B3D54AA0E099EE
