
Midmark M11 - v1.0.5
Total Cycles: 3733

Name: RITTER2
Sterilizer ID: V2313228
06/25/2024 12:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 90.9 F 0.0
2:00 92.0 F 0.5
4:00 107.3 F 2.2
6:00 145.9 F 4.8
8:00 170.9 F 7.9
10:00 184.5 F 11.2
12:00 202.4 F 4.8
14:00 212.3 F 6.4
16:00 222.7 F 10.2
18:00 230.8 F 9.4
20:00 236.5 F 10.8
22:00 235.1 F 8.4
24:00 247.4 F 14.6
26:00 258.5 F 20.7
28:00 268.8 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.8 F 29.9
2:00 271.6 F 29.1
3:00 272.5 F 30.1
4:00 272.4 F 29.5

Min 271.5 F 29.1
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 200.6 F 0.0
10:00 204.2 F 0.0
15:00 191.4 F 0.0
20:00 192.3 F 0.0
25:00 196.0 F 0.0
30:00 196.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:12
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:07:47
CYCLE COMPLETE

06/25/2024 01:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____