

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001763  
8:33 15/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:16  
PRESSURIZING 2:20  
STERILIZING 3:17  
136.4 C 223kPa 3:17  
136.3 C 223kPa 3:47  
136.4 C 224kPa 4:17  
136.4 C 224kPa 4:47  
136.4 C 223kPa 5:17  
136.3 C 223kPa 5:47  
136.2 C 222kPa 6:17  
136.2 C 219kPa 6:47  
136.2 C 219kPa 6:47  
VENTING 6:47  
AIR DRYING 7:39  
CYCLE COMPLETE 7:44

-----  
Digital Signature #  
F5DEC91E1CC7F7BA  
-----