
Midmark M11 - v1.0.0
Total Cycles: 4019

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2023 03:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.6 F 0.0
2:00 114.4 F 1.0
4:00 127.0 F 3.6
6:00 161.9 F 7.1
8:00 191.6 F 11.2
10:00 208.7 F 3.9
12:00 218.2 F 8.0
14:00 230.2 F 12.3
16:00 232.1 F 8.2
18:00 244.0 F 13.5
20:00 242.1 F 11.4
22:00 254.2 F 17.5
24:00 265.0 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.6 F 28.9
2:00 271.5 F 28.2
3:00 272.1 F 28.8
4:00 272.4 F 28.8

Min 271.2 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 217.1 F 0.0
10:00 211.9 F 0.0
15:00 213.6 F 0.0
20:00 231.5 F 0.0
25:00 231.3 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 25:54
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:04:18
CYCLE COMPLETE

11/14/2023 04:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____