

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1383

Name: RITTER 1  
Sterilizer ID: V2178198  
11/18/2021 10:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.2 F 0.0  
2:00 177.3 F 2.9  
4:00 189.6 F 7.1  
6:00 211.0 F 12.8  
8:00 229.7 F 7.4  
10:00 240.5 F 12.6  
12:00 238.2 F 10.0  
14:00 235.9 F 8.3  
16:00 249.9 F 15.2  
18:00 262.2 F 22.2  
20:00 272.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.5 F 29.0  
2:00 271.6 F 28.3  
3:00 272.0 F 28.9  
4:00 272.7 F 28.9  
  
Min 271.3 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.3 F 0.0  
5:00 234.5 F 0.0  
10:00 209.2 F 0.0  
15:00 225.2 F 0.0  
20:00 219.3 F 0.0  
25:00 213.0 F 0.0  
30:00 211.8 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 20:19  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:06  
CYCLE COMPLETE

11/18/2021 11:39 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02