
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002372
9:41 13/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:38
STERILIZING 3:48
136.4 C 224kPa 3:48
136.3 C 224kPa 4:18
136.4 C 222kPa 4:48
135.9 C 221kPa 5:18
136.3 C 222kPa 5:48
136.3 C 224kPa 6:18
136.3 C 224kPa 6:48
136.4 C 225kPa 7:18
136.4 C 225kPa 7:18
VENTING 7:18
AIR DRYING 8:09
CYCLE COMPLETE 37:29

Digital Signature #
32C67736A2DD15E4
