
Midmark M11 - v1.0.0
Total Cycles: 1252

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2021 06:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 84.1 F 0.0
4:00 81.9 F 0.0
6:00 109.3 F 0.1
8:00 142.0 F 0.1
10:00 173.4 F 0.2
12:00 192.7 F 0.5
14:00 219.0 F 3.7
16:00 226.9 F 6.4
18:00 233.6 F 9.5
20:00 240.6 F 13.1
22:00 236.8 F 9.5
24:00 240.3 F 10.2
26:00 244.5 F 12.4
28:00 241.7 F 10.8
30:00 253.8 F 16.9
32:00 264.3 F 23.1
34:00 271.5 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.4 F 28.3
2:00 272.7 F 28.8
3:00 271.9 F 28.6
4:00 272.1 F 28.3

Min 271.5 F 28.1
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.8 F 0.0
5:00 243.6 F 0.0
10:00 222.1 F 0.0
15:00 217.6 F 0.0
20:00 219.2 F 0.0
25:00 220.0 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 34:46
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:12:54
CYCLE COMPLETE

10/12/2021 08:06 AM
ACCEPTED: X
REJECTED:
Operator: U01