
Midmark M11 - v1.0.5
Total Cycles: 1948

Name: RITTER2
Sterilizer ID: V2313228
01/19/2023 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.6 F 0.0
2:00 152.8 F 1.5
4:00 166.3 F 4.6
6:00 173.9 F 8.8
8:00 197.6 F 10.0
10:00 214.3 F 7.4
12:00 233.1 F 12.4
14:00 232.0 F 9.4
16:00 237.6 F 9.5
18:00 246.3 F 14.1
20:00 258.1 F 20.7
22:00 269.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 271.5 F 29.5
2:00 272.7 F 30.2
3:00 272.7 F 29.8
4:00 271.8 F 29.4

Min 271.3 F 29.2
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 230.6 F 0.0
10:00 229.9 F 0.0
15:00 225.5 F 0.0
20:00 232.1 F 0.0
25:00 226.9 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 23:01
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

01/19/2023 10:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____