

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5493

Name:  
Sterilizer ID: V2178198  
04/10/2025 09:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.3 F 0.0  
2:00 71.3 F 0.4  
4:00 83.6 F 1.8  
6:00 117.8 F 4.2  
8:00 149.2 F 6.9  
10:00 171.1 F 10.3  
12:00 196.1 F 7.9  
14:00 202.3 F 3.6  
16:00 213.6 F 6.7  
18:00 224.3 F 10.2  
20:00 233.8 F 9.2  
22:00 236.5 F 9.7  
24:00 239.9 F 9.8  
26:00 245.3 F 12.7  
28:00 256.2 F 18.4  
30:00 266.2 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.5 F 28.9  
2:00 271.5 F 28.1  
3:00 272.4 F 29.0  
4:00 271.9 F 28.5  
  
Min 271.3 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.3 F 0.0  
5:00 198.5 F 0.0  
10:00 181.5 F 0.0  
15:00 183.6 F 0.0  
20:00 197.6 F 0.0  
25:00 216.3 F 0.0  
30:00 236.7 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 31:45  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:10:18  
CYCLE COMPLETE

04/10/2025 10:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_