
Midmark M11 - v1.0.0
Total Cycles: 4076

Name: RITTER 1
Sterilizer ID: V2178198
12/04/2023 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.5 F 0.0
2:00 69.0 F 0.2
4:00 78.8 F 1.0
6:00 113.4 F 2.9
8:00 144.0 F 5.2
10:00 167.5 F 8.0
12:00 215.9 F 11.1
14:00 217.1 F 7.6
16:00 212.7 F 5.7
18:00 221.1 F 8.9
20:00 232.1 F 12.5
22:00 232.7 F 7.9
24:00 242.1 F 12.6
26:00 238.9 F 10.0
28:00 250.5 F 15.6
30:00 260.4 F 21.3
32:00 270.0 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.7 F 29.2
2:00 272.7 F 29.3
3:00 272.7 F 29.3
4:00 272.1 F 29.4

Min 270.8 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 232.3 F 0.0
10:00 194.4 F 0.0
15:00 175.2 F 0.0
20:00 173.6 F 0.0
25:00 173.1 F 0.0
30:00 173.4 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:51
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:11:19
CYCLE COMPLETE

12/04/2023 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____