
Midmark M11 - v1.0.0
Total Cycles: 3776

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2023 03:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.4 F 0.0
2:00 173.8 F 2.2
4:00 189.7 F 5.9
6:00 207.1 F 10.2
8:00 230.9 F 7.3
10:00 228.1 F 8.9
12:00 241.0 F 11.3
14:00 241.1 F 11.5
16:00 239.5 F 10.1
18:00 252.3 F 16.3
20:00 263.1 F 22.6
22:00 272.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.7
2:00 271.5 F 28.1
3:00 271.9 F 28.7
4:00 272.1 F 28.6

Min 271.2 F 28.0
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 223.3 F 0.0
10:00 230.0 F 0.0
15:00 230.5 F 0.0
20:00 228.0 F 0.0
25:00 223.3 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 22:18
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:01:54
CYCLE COMPLETE

09/14/2023 04:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____