

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4450

Name: RITTER2  
Sterilizer ID: V2313228  
03/24/2025 04:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.6 F 0.0  
2:00 157.7 F 2.8  
4:00 174.7 F 7.1  
6:00 198.9 F 12.0  
8:00 220.5 F 6.8  
10:00 232.5 F 11.8  
12:00 234.5 F 9.2  
14:00 238.4 F 9.9  
16:00 246.7 F 14.3  
18:00 258.8 F 20.9  
20:00 269.8 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.4  
1:00 271.5 F 29.3  
2:00 271.6 F 29.3  
3:00 271.9 F 29.5  
4:00 272.1 F 29.8  
  
Min 271.1 F 29.3  
Max 272.4 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.9 F 0.0  
5:00 226.8 F 0.0  
10:00 235.8 F 0.0  
15:00 228.0 F 0.0  
20:00 222.7 F 0.0  
25:00 215.5 F 0.0  
30:00 187.5 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 20:52  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:00:03  
CYCLE COMPLETE

03/24/2025 05:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_