
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 006213
15:10 30/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:42
STERILIZING 3:48
136.4 C 224kPa 3:48
136.4 C 225kPa 4:18
136.2 C 222kPa 4:48
136.4 C 225kPa 5:18
136.4 C 224kPa 5:48
136.4 C 224kPa 6:18
136.4 C 224kPa 6:48
136.2 C 224kPa 7:18
136.2 C 224kPa 7:18
VENTING 7:18
AIR DRYING 8:07
CYCLE COMPLETE 27:00

Digital Signature #
07EF20A466442703
