
Midmark M11 - v1.0.0
Total Cycles: 3204

Name: RITTER 1
Sterilizer ID: V2178198
05/04/2023 01:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.2 F 0.0
2:00 182.6 F 2.8
4:00 193.6 F 7.2
6:00 215.2 F 12.5
8:00 229.2 F 7.7
10:00 241.6 F 13.3
12:00 240.0 F 11.1
14:00 238.5 F 9.7
16:00 252.8 F 16.8
18:00 265.0 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.8 F 29.2
2:00 271.7 F 28.4
3:00 272.2 F 29.2
4:00 272.5 F 29.0

Min 271.4 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 224.2 F 0.0
10:00 216.9 F 0.0
15:00 224.3 F 0.0
20:00 229.9 F 0.0
25:00 227.5 F 0.0
30:00 216.9 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 19:48
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

05/04/2023 02:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____