

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4557

Name: RITTER 1  
Sterilizer ID: V2178198  
05/15/2024 12:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 128.0 F 0.0  
2:00 134.3 F 1.6  
4:00 150.5 F 4.5  
6:00 172.0 F 7.8  
8:00 189.5 F 11.2  
10:00 212.5 F 3.4  
12:00 217.9 F 7.0  
14:00 226.1 F 11.0  
16:00 230.6 F 7.0  
18:00 240.3 F 11.8  
20:00 237.9 F 9.7  
22:00 249.9 F 15.4  
24:00 260.3 F 21.3  
26:00 270.3 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.4  
1:00 272.3 F 29.1  
2:00 271.5 F 28.4  
3:00 271.7 F 29.0  
4:00 273.0 F 29.5

Min 271.0 F 28.3  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
5:00 194.9 F 0.0  
10:00 189.5 F 0.0  
15:00 189.4 F 0.0  
20:00 203.3 F 0.0  
25:00 213.7 F 0.0  
30:00 207.9 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 26:54  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:05:47  
CYCLE COMPLETE

05/15/2024 01:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_