
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002325
15:35 21/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:57
STERILIZING 3:01
136.4 C 225kPa 3:01
136.3 C 225kPa 3:31
136.2 C 222kPa 4:01
136.4 C 226kPa 4:31
136.4 C 226kPa 5:01
136.2 C 223kPa 5:31
136.4 C 222kPa 6:01
136.4 C 225kPa 6:31
136.4 C 225kPa 6:31
VENTING 6:31
AIR DRYING 7:23
CYCLE COMPLETE 16:50

Digital Signature #
860E358DC4BDA740
