

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3704

Name: RITTER 1  
Sterilizer ID: V2178198  
08/30/2023 10:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
2:00 175.3 F 3.2  
4:00 190.6 F 7.7  
6:00 220.7 F 13.0  
8:00 224.6 F 7.9  
10:00 237.2 F 13.4  
12:00 240.0 F 10.8  
14:00 236.5 F 9.0  
16:00 251.3 F 15.9  
18:00 263.3 F 22.8  
20:00 272.0 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.7 F 29.4  
2:00 272.1 F 29.1  
3:00 271.9 F 28.7  
4:00 272.8 F 29.2  
  
Min 270.8 F 28.1  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.1 F 0.0  
5:00 233.6 F 0.0  
10:00 212.1 F 0.0  
15:00 214.7 F 0.0  
20:00 209.5 F 0.0  
25:00 210.3 F 0.0  
30:00 204.3 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 20:07  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:58:60  
CYCLE COMPLETE

08/30/2023 11:36 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_