
Midmark M11 - v1.0.0
Total Cycles: 4406

Name: RITTER 1
Sterilizer ID: V2178198
03/26/2024 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.9 F 0.0
2:00 87.6 F 0.5
4:00 99.2 F 2.0
6:00 120.5 F 4.2
8:00 156.8 F 6.9
10:00 197.9 F 10.2
12:00 218.6 F 7.9
14:00 215.2 F 4.9
16:00 226.1 F 8.3
18:00 234.5 F 12.1
20:00 231.7 F 7.7
22:00 241.8 F 12.6
24:00 240.0 F 10.2
26:00 251.3 F 15.7
28:00 261.1 F 21.5
30:00 270.5 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 271.6 F 28.7
2:00 271.6 F 28.6
3:00 271.9 F 28.7
4:00 272.5 F 28.9

Min 271.5 F 28.6
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.1 F 0.0
5:00 235.6 F 0.0
10:00 220.4 F 0.0
15:00 215.7 F 0.0
20:00 205.4 F 0.0
25:00 201.4 F 0.0
30:00 211.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 31:25
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:10:08
CYCLE COMPLETE

03/26/2024 02:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____