
Midmark M11 - v1.0.5
Total Cycles: 451

Name: RITTER2
Sterilizer ID: V2313228
10/14/2021 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.3 F 0.0
2:00 180.2 F 4.0
4:00 202.3 F 9.8
6:00 224.9 F 6.3
8:00 237.4 F 11.7
10:00 237.0 F 10.0
12:00 236.9 F 9.1
14:00 251.5 F 16.7
16:00 263.9 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.5 F 29.8
2:00 271.6 F 29.0
3:00 272.5 F 30.0
4:00 272.0 F 29.3

Min 271.4 F 28.9
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 234.8 F 0.0
10:00 218.4 F 0.0
15:00 199.2 F 0.0
20:00 195.6 F 0.0
25:00 181.9 F 0.0
30:00 175.2 F 0.0

Drying Complete

FILLING: 2:38
HEATING: 17:58
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:42
CYCLE COMPLETE

10/14/2021 11:47 AM
ACCEPTED: X
REJECTED:
Operator: U02