

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3753

Name: RITTER2  
Sterilizer ID: V2313228  
07/03/2024 02:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 90.9 F 0.0  
2:00 91.6 F 0.3  
4:00 122.9 F 2.4  
6:00 153.7 F 5.3  
8:00 175.9 F 8.7  
10:00 192.1 F 12.6  
12:00 212.5 F 5.8  
14:00 222.6 F 10.2  
16:00 233.3 F 8.3  
18:00 238.2 F 11.7  
20:00 236.5 F 9.7  
22:00 249.8 F 15.9  
24:00 261.0 F 22.3  
26:00 271.3 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.8  
1:00 272.4 F 29.7  
2:00 273.0 F 30.0  
3:00 273.0 F 30.3  
4:00 272.2 F 30.2

Min 271.8 F 29.7  
Max 273.3 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
5:00 220.8 F 0.0  
10:00 204.8 F 0.0  
15:00 218.5 F 0.0  
20:00 222.8 F 0.0  
25:00 234.7 F 0.0  
30:00 231.0 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 27:24  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:05:48  
CYCLE COMPLETE

07/03/2024 03:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_