
Midmark M11 - v1.0.0
Total Cycles: 1078

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2021 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.1 F 0.0
2:00 137.1 F 0.4
4:00 151.6 F 2.0
6:00 184.8 F 5.0
8:00 197.9 F 8.9
10:00 215.1 F 13.0
12:00 225.2 F 6.1
14:00 233.3 F 10.4
16:00 237.9 F 8.4
18:00 242.9 F 11.8
20:00 238.8 F 9.8
22:00 251.3 F 15.9
24:00 262.0 F 22.1
26:00 272.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.5 F 29.1
2:00 272.2 F 28.6
3:00 272.0 F 28.8
4:00 272.7 F 29.0

Min 271.6 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 201.5 F 0.0
10:00 205.4 F 0.0
15:00 194.3 F 0.0
20:00 191.4 F 0.0
25:00 190.7 F 0.0
30:00 203.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:25
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:04:48
CYCLE COMPLETE

08/09/2021 10:57 AM
ACCEPTED: X
REJECTED:
Operator: U01