
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001580
10:42 11/12/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:13
PRESSURIZING 6:18
STERILIZING 7:54
136.2 C 220kPa 7:54
136.1 C 220kPa 8:24
136.0 C 218kPa 8:54
136.4 C 222kPa 9:24
136.3 C 221kPa 9:54
135.9 C 220kPa 10:24
136.3 C 217kPa 10:54
136.5 C 223kPa 11:24
VENTING 11:24
AIR DRYING 12:24
CYCLE COMPLETE 42:24

Digital Signature #
83A635A843D9AE86
