
Midmark M11 - v1.0.5
Total Cycles: 432

Name: RITTER2
Sterilizer ID: V2313228
10/08/2021 08:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.1 F 0.0
2:00 191.0 F 3.2
4:00 201.5 F 8.6
6:00 218.7 F 10.6
8:00 229.1 F 9.7
10:00 235.9 F 8.9
12:00 243.2 F 11.6
14:00 246.2 F 13.9
16:00 259.0 F 20.9
18:00 270.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.3 F 29.2
2:00 272.8 F 29.6
3:00 272.4 F 29.8
4:00 271.9 F 29.0

Min 271.8 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 190.0 F 0.0
10:00 190.6 F 0.0
15:00 190.7 F 0.0
20:00 184.4 F 0.0
Dry Cycle Aborted

FILLING: 2:16
HEATING: 18:48
STERILIZING: 4:00
VENTING: 3:10
DRYING: 24:36
TOTAL CYCLE: 00:52:50

10/08/2021 09:49 AM
ACCEPTED: X
REJECTED:
Operator: U02