
Midmark M11 - v1.0.5
Total Cycles: 3852

Name: RITTER2
Sterilizer ID: V2313228
08/13/2024 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.2 F 0.0
2:00 125.0 F 1.0
4:00 129.6 F 3.1
6:00 171.6 F 6.5
8:00 189.0 F 10.5
10:00 221.6 F 6.0
12:00 221.3 F 8.5
14:00 231.2 F 13.0
16:00 234.6 F 9.3
18:00 240.5 F 10.6
20:00 244.4 F 12.8
22:00 255.8 F 18.9
24:00 266.4 F 25.4
26:00 272.5 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 271.5 F 29.5
2:00 271.4 F 29.5
3:00 271.6 F 29.4
4:00 272.0 F 29.5

Min 271.3 F 29.4
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 169.5 F 0.0
5:00 231.1 F 0.0
10:00 183.7 F 0.0
15:00 179.7 F 0.0
20:00 176.5 F 0.0
25:00 176.0 F 0.0
30:00 176.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:22
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:04:55
CYCLE COMPLETE

08/13/2024 12:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____