

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000457  
14:35 01/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:40  
STERILIZING 3:56  
136.2 C 223kPa 3:56  
136.4 C 233kPa 4:26  
136.1 C 229kPa 4:56  
136.2 C 232kPa 5:26  
136.2 C 232kPa 5:56  
136.2 C 232kPa 6:26  
136.4 C 233kPa 6:56  
136.4 C 233kPa 7:26  
136.4 C 233kPa 7:26  
VENTING 7:26  
AIR DRYING 8:23  
CYCLE COMPLETE 8:24

-----  
Digital Signature #  
6B51F561851B5008  
-----